



**TYPE APPROVAL CERTIFICATE**  
No. **MAC277419XG**

**This is to certify** that the product below is found to be in compliance with the applicable requirement of the RINA type approval system.

<i>Description</i>	<b>Multigas Monitoring System</b>
<i>Type</i>	<b>G7000</b>
<i>Applicant</i>	<b>Green Instruments A/S Erhvervsparken 29 DK- 9700 Brønderslev DENMARK</b>
<i>Manufacturer</i>	<b>Green Instruments A/S</b>
<i>Place of manufacture</i>	<b>Erhvervsparken 29 DK- 9700 Brønderslev DENMARK</b>
<i>Reference standards</i>	<b>NOx Technical Code 2008 ; MEPC.259(68) &amp; MEPC.340(77); RINA Rules for the Classification of Ships - Part C - Machinery, Systems and fire protection - Ch.3, Sect.6, Tab.1.</b>

Issued in **HAMBURG** on **September 26, 2022**. This Certificate is valid until **September 23, 2024**

\_\_\_\_\_  
**RINA Services S.p.A.**  
**Giuseppe Russo**

This certificate consists of this page and 1 enclosure.

## TYPE APPROVAL CERTIFICATE

No. MAC277419XG

Enclosure - Page 1 of 1

G7000

### Product Description:

This system is a continuous emission monitoring system. The modular design of the system makes it possible to monitor different gases according to customer specification. It can monitor the SO<sub>2</sub> and CO<sub>2</sub> concentrations in exhaust gases. It provides an accurate measurement of SO<sub>2</sub> in ppm, CO<sub>2</sub> in percent as well as presents the SO<sub>2</sub>/CO<sub>2</sub> ratio.

### Technical Data:

The system consists of the following components:

#### Heated sample probe(s) with heated sample hose(s):

- Up to 5 sample probes each with heated hoses leading to the monitoring cabinet
- Sample flow: 0.33-1.67 l/min
- Exhaust gas temperature 0-500°C
- Ambient temperature: 5°C to 55°C

#### Monitoring Cabinet with HMI control panel:

- Enclosure: IP55
- Power Supply: 230V AC 50/60 Hz
- Communication: Modbus TCP/IP

#### Gas Analyzer:

- Type: ABB Uras 26
- Measurement Principle: NDIR
- Gasses: Measures CO<sub>2</sub> and SO<sub>2</sub>. Other gasses available upon request.
- Measuring range: CO<sub>2</sub>: 0-10%; freely configurable from 0-20%  
SO<sub>2</sub>: 0-200 ppm; freely configurable from 0-1000ppm
- Accuracy/Linearity:  $\leq \pm 2\%$  of reading, or  $\leq \pm 0,3\%$  of full scale whichever is larger
- Precision/Repeatability:  $\leq \pm 1\%$  of full scale above 100ppm or  $\leq \pm 2\%$  of full scale below 100 ppm
- Noise:  $\leq \pm 2\%$  of full scale
- Zero & Span Drift:  $\leq \pm 2\%$  of full scale
- Calibration: Zero Calibration: Automatic using compressed air  
Span Calibration: Automatic using built in optical reference filters. Possibility for verification with test gasses.

#### Reference Documents:

P&I + Installation Layout + Component + Electrical Drawings filed under RINA dwg no. HMMC-14803

Test Reports + Installation & Operation Manual filed under RINA dwg no. HMMC-14804

#### Application Fields:

The G7000 Multigas Monitoring System may be used on board ships and other units classed by the RINA for the continuous monitoring of different gases in exhaust gas ducts according to the requirements of the NO<sub>x</sub> Technical Code 2008 Appendix 3 and IMO Res. MEPC. 259(68) and MEPC.340(77) .

#### Acceptance condition:

Correct configuration and set up for each delivery to be tested during commissioning after installation.

The monitoring cabinet shall be mounted on vibration dampers according to the manufacturer instructions.

HAMBURG September 26, 2022

