



Certificate number: 22859/D0 BV

File number: AP 4004

Product code: 3795H

This certificate is not valid when presented without the full attached schedule composed of 7 sections

www.veristar.com

TYPE APPROVAL CERTIFICATE

This certificate is issued to

Green Instruments A/S

Brønderslev - DENMARK

for the type of product

OXYGEN GAS DETECTORS / ANALYSERS

Oxygen Analysers: G36a and G36p

Requirements:

BUREAU VERITAS Rules for the Classification of Steel Ships.

EC Code : 31B

This certificate is issued to attest that Bureau Veritas Marine & Offshore did undertake the relevant approval procedures for the product identified above which was found to comply with the relevant requirements mentioned above.

This certificate will expire on: 14 Oct 2030

For Bureau Veritas Marine & Offshore,

At BV FREDERICIA, on 14 Oct 2025,

Jesper JENSEN

This certificate was created electronically and is valid without signature



This certificate remains valid until the date stated above, unless cancelled or revoked, provided the conditions indicated in the subsequent page(s) are complied with and the product remains satisfactory in service. This certificate will not be valid if the applicant makes any changes or modifications to the approved product, which have not been notified to, and agreed in writing with Bureau Veritas Marine & Offshore. Should the specified regulations or standards be amended during the validity of this certificate, the product(s) is/are to be re-approved prior to it/they being placed on board vessels to which the amended regulations or standards apply. This certificate is issued within the scope of the General Conditions of Bureau Veritas Marine & Offshore available on the internet site www.veristar.com. Any Person not a party to the contract pursuant to which this document is delivered may not assert a claim against Bureau Veritas Marine & Offshore for any liability arising out of errors or omissions which may be contained in said document, or for errors of judgement, fault or negligence committed by personnel of the Society or of its Agents in establishment or issuance of this document, and in connection with any activities for which it may provide.

The electronic version is available at: <https://www.veristarm.com/veristarnb/jsp/viewPublicPdfTypepec.jsp?id=bkkj3uggph>

BV Mod. Ad.E 530 June 2017

This certificate consists of 3 page(s)

THE SCHEDULE OF APPROVAL

1. PRODUCT DESCRIPTION :

G36a and **G36p** are Oxygen Analysers based on a Heated Zirconia Sensor intended to measure the concentration of Oxygen in Nitrogen or Flue gas based Inert gas or in the atmosphere in engine room compartments or rooms for inert gas system or accommodation area.

G36 Systems comprise the following parts:

Main Unit / Analyser: Measuring Range: 0.0 to 20.9 % Oxygen
 Ambient Temperature: -15°C to 55°C (G36), 70°C (G36p)
 Power Supply:
 G36a: 90/230 VAC - 50/60 Hz, G36d: 24 VDC
 G36p: 24 VDC (Flush mounting)
 Analogue Output: 2 x 4-20 mA
 Output Relay: 4 x Volt Free Contact Relay (250 VAC - 5A)
 Alarm: High, High High O2 level (Free Configurable) and System-Fail
 Interface: Touch Screen 71x39 mm with trend graph display
 G36a and G36d: IP67
 G36p: IP55 (if panel mounted)

Oxygen Sensor: Heated Zirconia Sensor, mounted in sampling line
(SEN1 and SEN9 parts)

Software Release: V2.0x

2. DOCUMENTS AND DRAWINGS :

GREEN Instruments:

Manuals Oxygen Analyser : G36a No. 104631 and G36p No. 104991.

Electronic Drawings:

P0744 (5 pages) Rev. 1.1 dated March 26, 2008
 P0744B (5 pages) Rev. 2 dated April 28, 2008
 P0745 Rev. 1 dated February 14, 2008
 P0901 Rev. 1B dated March 26, 2009
 P0902 Rev. A2 dated March 02, 2009.

3. TEST REPORTS :

GREEN Instruments:

- In-house Performance Test G36 Oxygen Analyser according to EN 50104(2019), dated 04-03-2024.
 - Stability Test G36 Oxygen Analyser according to EN 50104(2019), dated 05-03-2024.

EKTOS:

- Marine tests report No. P17-0136, dated 28-12-2017.
 - Type approval test No. P24-0020-1 rev.2, dated 08-04-2024.

4. APPLICATION / LIMITATION :

- 4.1 - BUREAU VERITAS Rules for the Classification of Steel Ships.
- 4.2 - Approval also valid for ships to be granted with the notations: **AUT-UMS, AUT-CCS, AUT-PORT and AUT-IMS.**
- 4.3 - BUREAU VERITAS **EC Code: 31B.**
- 4.4 - The equipment fulfils the EMC requirements for an installation on the Bridge and Deck Zone.
- 4.5 - Not to be installed in hazardous area and / or in presence of flammable gases.
- 4.6 - Not to be installed in non weather protected area.
- 4.7 - Only Hardware and Firmware / Software successfully tested together in compliance with the rules as referred to in cover page, according to the declaration of the manufacturer are covered by this certificate.
- 4.8 - Any modification of the hardware, firmware or software having an impact on the product performance or functionality has to be validated with type testing.
- 4.9 - The equipment, once installed on board ship, is to be tested in accordance with the above referred Rules under the supervision of a Society's Surveyor.
- 4.10 - Equipment covered by this Type Approval certificate has been tested according to requirements of IACS UR E10 rev 9.

5. PRODUCTION SURVEY REQUIREMENTS :

- 5.1 - The G36a and G36p are to be supplied by **Green Instruments A/S** in compliance with the type described in this certificate.
- 5.2 - This type of product is within the category HBV of Bureau Veritas Rule Note NR320 and as such does not require a BV product certificate.
- 5.3 - **Green Instruments A/S** has to make the necessary arrangements to have its works recognised by Bureau Veritas in compliance with the requirements of NR320 for HBV products.
- 5.4 - For information, **Green Instruments A/S** has declared to Bureau Veritas the following production site(s):

Green Instruments A/S
Erhvervsparken 29
9700 Brønderslev
DENMAR

6. MARKING OF PRODUCT :

- Trade name.
- Date of manufacture and serial number.
- Equipment type or model identification under which it was type-tested.
- Alternatively, the marking may be presented on a display at equipment start-up.
- The title and version of each software element included in the installed software system shall be either marked or displayed on command on the equipment.
- When the marking and the title and version of the software are displayed only on the display, such information shall also be included in the equipment manual.

7. OTHERS :

- 7.1 - It is **Green Instruments A/S's** responsibility to inform shipbuilders or their sub-contractors of the proper methods of fitting, use and general maintenance of the approved equipment and the conditions of this approval.
- 7.2 - This certificate supersedes the Type Approval Certificate No. 22859/C0 BV issued by the Society.

*** END OF CERTIFICATE ***